

# Directive

9180.61

August 1, 2021

---

## OFFICIAL MOISTURE CALIBRATIONS FOR UNIFIED GRAIN MOISTURE ALGORITHM (UGMA) COMPATIBLE METERS

### Contents

1. PURPOSE .....	2
2. BACKGROUND .....	2
3. REPLACEMENT HIGHLIGHTS .....	3
4. PROCEDURES .....	3
5. QUESTIONS .....	3
Attachment 1: Grain types as displayed on the moisture meters with current calibration version identifiers. ....	4
Attachment 2: Official Calibration Moisture and Temperature Ranges .....	7

The U.S. Department of Agriculture (USDA) prohibits discrimination in its programs on the basis of race, color, national origin, sex, religion, age, disability, political beliefs, and marital or familial status. (Not all prohibited bases apply to all programs.) Persons with disabilities who require alternate means for communication of program information (Braille, large print, audiotape, etc.) should contact USDA's TARGET Center at (202) 720-2600 (voice and TDD).

To file a complaint, write to the USDA, Office of Civil Rights, Room 326-W, 1400 Independence Avenue, SW, Washington, DC 20250-9410, or call (202) 720-5964 (voice and TDD). USDA is an equal employment opportunity employer.

## 1. PURPOSE

This directive defines the official calibrations for use with the DICKEY-john GAC 2500-UGMA, Perten AM 5200-A, and FOSS GAC 2500-C moisture meters used by the Agricultural Marketing Service (AMS) and its system of Official Inspection Providers for official grain moisture measurement.

## 2. BACKGROUND

This directive specifies updated official UGMA-compatible moisture meter calibrations for Corn (Perten), Long Grain Milled Rice (All Models), Medium Grain Milled Rice (All Models), Sunflower Seed (oil-type) ( DICKEY-john/FOSS), Sunflower Seed (Confectionary) (All Models), Long Grain Brown Parboiled Rice (Perten), Second Head Milled Parboiled Rice (All Models), Long Grain Milled Parboiled Rice (All Models), Medium Grain Milled Parboiled Rice (All Models), Brewers Milled Parboiled Rice (All Models), Baby Lima Beans (DICKEY-john/FOSS), Garbanzo Beans (DICKEY-john/FOSS), Small Red Beans (DICKEY-john/FOSS), and Yelloweye Beans (DICKEY-john/FOSS) through which moisture results will be changed.

To maintain compatibility with state regulatory programs for moisture meters certified for test weight measurements, test weight coefficients were updated for Long Grain Milled Rice (DICKEY-john/FOSS), Medium Grain Milled Rice (DICKEY-john/FOSS), Baby Lima Beans (Perten), Garbanzo Beans (Perten), Small Red Beans (Perten), Yelloweye Beans (Perten). Test weight coefficients are combined with moisture coefficients into a single calibration file and the changes for these grains are for test weight only and will not affect moisture results unless otherwise noted above.

This directive also restates all other calibrations transmitted in the previous directive dated May 1, 2021.

AMS requires that its official moisture instruments conform to National Type Evaluation Program requirements so they can be used for commercial as well as official grain inspection. The NTEP requirements are stated in the National Conference on Weights and Measures: Publication 14 Grain Moisture Meters & Near Infrared Grain Analyzers ©2021.

AMS's calibration accuracy criteria, more restrictive than the NTEP requirements, mandates that the average difference between the results from a meter model and the USDA air oven method must not exceed +/- 0.15 percent moisture (for the most recent three years of data for the NTEP grains and for five years of data for the minor grains). The tolerance is increased slightly if the number of available samples is small. Tolerances for agreement between UGMA-Compatible meter models and AMS's master UGMA system are defined in CRITERIA FOR UNIFIED GRAIN MOISTURE ALGORITHM (UGMA) AND OFFICIAL SYSTEM COMPATIBILITY (available on the AMS website).

Attachment 1 lists the DICKEY-john GAC 2500-UGMA (FOSS GAC 2500-C) and Perten AM 5200-A calibration versions for the major grains (those grains that are also included in the NTEP) and for non-NTEP minor grains and commodities for official grain moisture measurements. The FOSS GAC 2500-C uses the same calibrations as the DICKEY-john GAC 2500-UGMA. Attachment 2 lists calibration moisture ranges and temperature ranges applicable to both meter models. The basic sample temperature range across all moisture contents and all grains is 0 to 45 °C (32 to 113 °F). The lower temperature limit is extended to -18 °C (-0.4 °F) for moisture levels below the limits indicated for each grain type in Attachment 2. Moisture readings are automatically inhibited when sample temperatures are outside the specified ranges.

### **3. REPLACEMENT HIGHLIGHTS**

This directive replaces Directive 9180.61, Official Moisture Meter Calibrations, dated, May 1, 2021.

### **4. PROCEDURES**

Official calibration files and installation instructions for both the DICKEY-john GAC 2500-UGMA and the Perten AM 5200-A are available for download at Moisture Equipment (<https://www.ams.usda.gov/services/fgis/standardization/moisture-equipment>). Official calibrations and updates to calibrations must be installed on the FOSS GAC 2500-C remotely through the FossManager network software. Please contact FOSS at (952) 974-9892 ext. 3 to request the latest calibration be updated to your GAC 2500-C. All official inspection laboratories shall use these official calibrations for the DICKEY-john GAC 2500-UGMA, Perten AM 5200-A, and FOSS GAC 2500-C meters for all grains and commodities for official inspections performed on and after August 1, 2021.

### **5. QUESTIONS**

Direct any questions concerning this directive to the Inspection Instrumentation Branch at (816) 702-3800.

Attachments

**Attachment 1: Grain types as displayed on the moisture meters with current calibration version identifiers.**

<b>Grain/Commodity Name</b>	<b>Displayed Grain Type</b>	<b>Perten AM 5200-A Calibration Version</b>	<b>DICKEY-john GAC 2500-UGMA Calibration Version</b>	<b>Calibration Effective Date*</b>
<b>NTEP GRAINS</b>				
Barley, Six-rowed	Barley 6Row	BAR 200417	20200415N	May 1, 2020
Barley, Two-rowed	Barley 2Row	BAR 200417	20200415N	May 1, 2020
Corn	Corn	CORN 210414	20200415N	Aug. 1, 2020 (D) Aug. 1, 2021 (P)
Oats	Oats	OATS 190415	20200415N	May 1, 2020 (D) May 1, 2019 (P)
Rice, Long Grain Rough	Rice LG Ruf	LGRR 170414	20200415N	Aug. 1, 2020 (D) Aug. 1, 2017 (P)
Rice, Medium Grain Rough	Rice MG Ruf	MGRR 190415	20190415N	Aug. 1, 2019
Sorghum	Sorghum	SORG 200417	20200415N	Aug. 1, 2020
Soybeans	Soybeans	SOY 190415	20190415N	Aug. 1, 2021 (D) Aug. 1, 2019 (P)
Sunflower seed (Oil-Type)	Sunflower	SFS 200417	20210414N	Aug. 1, 2021 (D) Aug. 1, 2020 (P)
Wheat, Durum	Durum	DURUM 180417	20170417N	May 1, 2019 (D) May 1, 2018 (P)
Wheat, Hard Red Spring	Wheat HRS	WHTH 190415	20200415N	May 1, 2020 (D) May 1, 2019 (P)
Wheat, Hard Red Winter	Wheat HRW	WHTH 190415	20200415N	May 1, 2020 (D) May 1, 2019 (P)
Wheat, Hard White	Wheat HW	WHTH 190415	20200415N	May 1, 2020 (D) May 1, 2019 (P)
Wheat, Soft Red Winter	Wheat SRW	WHTS 200417	20200415N	May 1, 2020
Wheat, Soft White	Wheat SW	WHTS 200417	20200415N	May 1, 2020
<b>NON-NTEP GRAINS</b>				
Beans, Baby Lima	BN Baby Lima	BEAN2 210414	20210414	Aug. 1, 2021
Beans, Black	BN Black	BEAN3 200417	20130501	May 1, 2013 (D) Aug. 1, 2020 (P)
Beans, Blackeye	BN Blackeye	BEAN1 200417	20130501	May 1, 2013 (D) Aug. 1, 2020 (P)
Beans, Cranberry	BN Cranberry	BEAN1 200417	20130501	May 1, 2013 (D) Aug. 1, 2020 (P)
Beans, Kidney	BN Kidney	BEAN5 200417	20180414	Aug. 1, 2018 (D) Aug. 1, 2020 (P)
Beans, Garbanzo	BN Garbanzo	BEAN2 210414	20210414	Aug. 1, 2021

**\*Note: Calibration effective dates with a (D) apply to the DICKEY-john GAC 2500-UGMA (FOSS GAC 2500-C), dates with a (P) apply to the Perten AM 5200-A, and dates without a ( ) apply to both models.**

<b>Grain/Commodity Name</b>	<b>Displayed Grain Type</b>	<b>Perten AM 5200-A Calibration Version</b>	<b>DICKEY-john GAC 2500-UGMA Calibration Version</b>	<b>Calibration Effective Date*</b>
Beans, Great Northern	BN Grt Nrthn	BEAN3 200417	20130501	May 1, 2013 (D) Aug. 1, 2020 (P)
Beans, Large Lima	BN Lrg Lima	BEAN3 200417	20130501	May 1, 2013 (D) Aug. 1, 2020 (P)
Beans, Pea	BN Pea	BEAN4 200417	20130501	May 1, 2013 (D) Aug. 1, 2020 (P)
Beans, Pink	BN Pink	BEAN1 200417	20130501	May 1, 2013 (D) Aug. 1, 2020 (P)
Beans, Pinto	BN Pinto	BEAN1 200417	20170417	Aug. 1, 2017 (D) Aug. 1, 2020 (P)
Beans, Small Red	BN Small Red	BEAN2 210414	20210414	Aug. 1, 2021
Beans, Small White	BN Small Wht	BEAN4 200417	20130501	May 1, 2013 (D) Aug. 1, 2020 (P)
Beans, Yelloweye	BN Yelloweye	BEAN2 210414	20210414	Aug. 1, 2021
Canola	Canola	CANOLA 190415	20190415	Aug. 1, 2019
Flaxseed	Flaxseed	FLAX 140418	20130501	May 1, 2013 (D) Aug. 1, 2014 (P)
Lentils	Lentils	BEAN5 200417	20180414	Aug. 1, 2018 (D) Aug. 1, 2020 (P)
Mustard seed, Brown	Mustard Brn	MUSTARD 130415	20130501	May 1, 2013
Mustard seed, Oriental	Mustard O	OMUSTARD 130415	20130501	May 1, 2013
Mustard seed, Yellow	Mustard Y	MUSTARD 130415	20130501	May 1, 2013
Peas, Mottled	Peas Mottled	PEAS 200417	20200415	Aug. 1, 2020
Peas, Smooth Dry	Peas Smooth	PEAS 200417	20200415	Aug. 1, 2020
Peas, Split	Peas Split	BEAN1 200417	20130501	May 1, 2013 (D) Aug. 1, 2020 (P)
Peas, Wrinkled Dry	Peas Wrinkl'd	PEAS 200417	20200415	Aug. 1, 2020
Popcorn	Popcorn	POPCORN 150420	20150420	Aug. 1, 2015
Rapeseed	Rapeseed	CANOLA 190415	20190415	Aug. 1, 2019
Rice, Brewers Milled	Rice BMR	SMRICE 140418	20140418	Aug. 1, 2014
Rice, Brewers Milled Parboiled	Rice BMRP	PBRICE 210414	20210414	Aug. 1, 2021
Rice, Long Grain Brown	Rice LGBR	BRRICE 140418	20140418	Aug. 1, 2014
Rice, Long Grain Brown Parboiled	Rice LGBRP	PBRICE 210414	20130501	May 1, 2013(D) Aug. 1, 2021(P)

**\*Note: Calibration effective dates with a (D) apply to the Dickey-john GAC 2500-UGMA (FOSS GAC 2500-C), dates with a (P) apply to the Perten AM 5200-A, and dates without a ( ) apply to both models.**

<b>Grain/Commodity Name</b>	<b>Displayed Grain Type</b>	<b>Perten AM 5200-A Calibration Version</b>	<b>DICKEY-john GAC 2500-UGMA Calibration Version</b>	<b>Calibration Effective Date*</b>
Rice, Long Grain Milled	Rice LGMR	LMRICE 210414	20210414	Aug. 1, 2021
Rice, Long Grain Milled Parboiled	Rice LGMRP	PBRICE 210414	20210414	Aug. 1, 2021
Rice, Medium Grain Brown	Rice MGBR	BRRICE 140418	20140418	Aug. 1, 2014
Rice, Medium Grain Milled	Rice MGMR	LMRICE 210414	20210414	Aug. 1, 2021
Rice, Medium Grain Milled Parboiled	Rice MGMRP	PBRICE 210414	20210414	Aug. 1, 2021
Rice, Screenings Milled	Rice SMR	SMRICE 140418	20140418	Aug. 1, 2014
Rice, Second Head Milled	Rice SHMR	SMRICE 140418	20140418	Aug. 1, 2014
Rice, Second Head Milled Parboiled	Rice SHMRP	PBRICE 210414	20210414	Aug. 1, 2021
Rice, Short Grain Brown	Rice SGBR	BRRICE 140418	20140418	Aug. 1, 2014
Rice, Short Grain Milled	Rice SGMR	SMRICE 140418	20140418	Aug. 1, 2014
Rice, Short Grain Rough	Rice SG Ruf	SGRR 190415	20190415	Aug. 1, 2019
Rye	Rye	RYE 140418	20130501	May 1, 2013 (D) May 1, 2014 (P)
Safflower seed	Safflower	SAFFL 140418	20130501	May 1, 2013 (D) Aug. 1, 2014 (P)
Sunflower seed (Confectionary)	Sunflwr Conf	SFS 210414	20210414	Aug. 1, 2021
Triticale	Triticale	RYE 140418	20130501	May 1, 2013 (D) May 1, 2014 (P)
<b>NON-STANDARDIZED GRAINS</b>				
Barley, Hulless	Barley Hless	BAR 200417	20200415	May 1, 2020
Buckwheat	Buckwheat	BKWHT 140418	20140418	Aug. 1, 2014
Buckwheat Groats	Buckwheat Gr	BKWHT 140418	20140418	Aug. 1, 2014
Oats, Hulless	Oats Hless	DURUM 150420	20150420	May 1, 2015

**\*Note: Calibration effective dates with a (D) apply to the DICKEY-john GAC 2500-UGMA (FOSS GAC 2500-C), dates with a (P) apply to the Perten AM 5200-A, and dates without a ( ) apply to both models.**

## Attachment 2: Official Calibration Moisture and Temperature Ranges

Official calibration moisture ranges (outside of which warnings are displayed) and upper moisture limits for moisture measurements when sample temperatures are below 0 °C (32 °F). The upper sample temperature limit is 45 °C (113 °F) for all grain types.

Grain/Commodity Name	Calibration Moisture Range (%)	Temperature limit extended to -18°C (0 °F) for moisture less than __%
<b>NTEP GRAINS</b>		
Barley, Six-rowed	8-18	18
Barley, Two-rowed	8-18	18
Corn	6-46	19
Oats	8-16	13
Rice, Long Grain Rough	8-26	18
Rice, Medium Grain Rough	8-28	18
Sorghum	10-24	16
Soybeans	5-24	20
Sunflower seed (Oil-Type)	4-28	13
Wheat, Durum	4-30	19
Wheat, Hard Red Spring	6-22	19
Wheat, Hard Red Winter	6-22	19
Wheat, Hard White	6-22	19
Wheat, Soft Red Winter	7-24	19
Wheat, Soft White	7-24	19
<b>NON-NTEP GRAINS</b>		
Beans, Baby Lima	6-20	15
Beans, Black	8-20	15
Beans, Blackeye	8-20	15
Beans, Cranberry	8-20	15
Beans, Kidney	6-20	15
Beans, Garbanzo	6-20	15
Beans, Great Northern	8-20	15
Beans, Large Lima	8-20	15
Beans, Pea	8-22	15
Beans, Pink	8-20	15
Beans, Pinto	8-20	15
Beans, Small Red	6-20	15
Beans, Small White	8-22	15
Beans, Yelloweye	6-20	15

<b>Grain/Commodity Name</b>	<b>Calibration Moisture Range (%)</b>	<b>Temperature limit extended to -18°C (0 °F) for moisture less than __%</b>
Canola	4-14	18
Flaxseed	6-14	10
Lentils	6-20	15
Mustard Seed, Brown	6-12	9
Mustard Seed, Oriental	6-18	9
Mustard Seed, Yellow	6-12	9
Peas, Mottled	6-16	15
Peas, Smooth Dry	6-16	15
Peas, Split	8-20	15
Peas, Wrinkled Dry	6-16	15
Popcorn	10-30	19
Rapeseed	4-14	18
Rice, Brewers Milled	10-16	16
Rice, Brewers Milled Parboiled	10-14	13
Rice, Long Grain Brown	10-22	16
Rice, Long Grain Brown Parboiled	10-14	13
Rice, Long Grain Milled	10-18	16
Rice, Long Grain Milled Parboiled	10-14	13
Rice, Medium Grain Brown	10-22	16
Rice, Medium Grain Milled	10-18	16
Rice, Medium Grain Milled Parboiled	10-14	13
Rice, Screenings Milled	10-16	16
Rice, Second Head Milled	10-16	16
Rice, Second Head Milled Parboiled	10-14	13
Rice, Short Grain Brown	10-22	16
Rice, Short Grain Milled	10-16	16
Rice, Short Grain Rough	10-28	18
Rye	8-16	16
Safflower seed	4-14	12
Sunflower seed (Confectionary)	4-28	13
Triticale	8-16	16
<b>NON-STANDARDIZED GRAINS</b>		
Barley, Hulless	8-18	18
Buckwheat	10-19	13
Buckwheat Groats	10-19	13
Oats, Hulless	4-30	19