


## LAPORAN PRAKTIKUM

### TEKNIK DASAR : PENGGUNAAN PIPET, TIMBANGAN DAN PEMBUATAN LARUTAN

**Hari/Tanggal Praktikum** : Rabu/ 2 Oktober 2013  
**Nama Mahasiswa** : 1. Nita Andriani Lubis  
 2. Ade Sinaga

**Tabel b.2. Penggunaan otomatis, mohr dan spuit**

Percobaan	Hasil pengukuran 1 ml aquadest (g)					
	Pipet Otomatik		Pipet Mohr		Pipet Spuit	
	Nita	Ade	Nita	Ade	Nita	Ade
1	0.988	0.99	0.92	0.93	1.070	0.897
2	0.985	0.98	0.99	0.85	1.080	0.919
3	0.986	0.98	0.93	0.94	1.086	0.939
4	0.990	0.98	0.95	0.93	1.003	0.954
5	0.985	0.98	0.94	0.92	1.021	0.916
Rata-rata	0.986	0.982	0.946	0.914	1.052	0.925
Standar deviasi	0.002168	0.00447	0.02702	0.03647	0.037503	0.022011



**Grafik Hasil pengukuran 1 ml aquadest (g)**